



CSIR-Indian Institute of Petroleum

(Council of Scientific & Industrial Research)

P.O.I.I.P., MOHKAMPUR, DEHRADUN – 248005 (UK) **INDIA**

Ph.0135-2525762 & 2525945, Fax. 0135-2660072, 2660202-203

E-Mail- pmitra@iip.res.in, rawats@iip.res.in

Website : www.iip.res.in



Subject: Invitation of Expression of Interest (EOI) No : 02/2018-19

Indian Institute of Petroleum (IIP), Dehradun, and ISO 9001 Institute, is one of the leading constituent laboratories under Council of Scientific & Industrial Research (CSIR) engaged in R&D work in petroleum refining, natural gas and petro-chemicals and contributing towards creation of state of the art technology & products. CSIR-IIP has been working on several projects of natural importance independency and also in collaboration with well-known Indian foreign organizations.

EOI are hereby invited from reputed engineering firms for fabrication, installation and commissioning of the following:

Sl.No	Reference No.	Description
1.	PUR/1/18-19/EOI/21/SR/BFD/PO	GC Analyzer for Refinery Gases (RGA)
2.	PUR/1/18-19/EOI/22/NK/BFD/PO	Bomb Calorimeter
3.	PUR/1/18-19/EOI/23/AN/BFD/PO	Advanced Proximate Analyzer
4.	PUR/1/18-19/EOI/24/AN/BFD/PO	Differential Scanning Calorimeter (DSC)
5.	PUR/1/18-19/EOI/25/NK/BFD/PO	CHNS and O Elemental Analyzer
6.	PUR/2/18-19/EOI/17/SJ/ECO-CAMPUS	Disposal of Hazardous Waste

Last date of submission : 23rd May, 2018 by 2:30 PM

Date of opening : 23rd May, 2018 at 3:00 PM

Interested bidders may download the details from our Website: www.iip.res.in

(S.O. Stores & Purchase)

Subject: Invitation for Expression of Interest (EOI) for GC Analyzer for Refinery Gases (RGA)

CSIR-Indian Institute of Petroleum (IIP), Dehradun, an ISO 9001 Institute, is one of the leading constituent laboratories under council of Scientific & Industrial Research (CSIR) engaged in R&D work in petroleum refining, natural gas and petro-chemicals and contributing towards creation of state of the art technology & products. CSIR-IIP has been working on several projects of national importance independently and also in collaboration with well-known Indian organizations.

EOI are invited from reputed firms for supply, installation and demonstration of (i) Refinery Gas Analyzer unit at CSIR-IIP, Dehradun. Gas Chromatograph with FID and TCD for specific applications: Pyrolysis gaseous products and oxygenates. (H₂, N₂, O₂, CO, CO₂, moisture, HC, C₁-C₅+, oxygenates in single online injection and demonstrate for the same. The scope of the work includes supply, installation and commissioning of the unit(s) at IIP.

Firms having done similar nature of work can apply along with the documentary evidence for the work done in the past. The firms should also meet the other parameters as given below and are required to submit following information along with their applications:

- a) Name of the firm with their constitution/proprietorship/partnership detail, etc with the date of establishment/registration
- b) List of similar works successfully completed in the last ten years as above with testimonials from department concerned and the details of contact persons.
- c) The firm should not have incurred any loss in more than 2 years during the last 5 years ending 31st March 2018.
- d) List of work in hand giving nature of work, department, and cost, date of start and completion with present progress and the contact details of the clients.
- e) Balance sheet of the firm for previous two years (2016-17 & 2017-18) must be enclosed with the offer certified by chartered accountant evidencing turnover.
- f) Please submit the article of association with offer to know the standing of the firm

Offers against this EOI should be submitted in a sealed envelope duly super-scribing "**Expression of Interest (EOI) for GC Analyzer for Refinery Gases (RGA)**" to **The Director, Indian Institute of Petroleum, Mohkampur, Haridwar Road, Dehradun-248005**. Kindly send your confirmation along with complete profile and offer **positively by 23/05/2018 by 2.30 PM**. The offers shall be opened on the same day at 3:00 PM. Shortlisted firms shall be called for making a presentation at a later date.

If any information furnished by the applicant is found incorrect at a later stage, he shall be liable to be debarred from tendering/taking up of work in CSIR. CSIR-IIP reserves the right to verify the particulars furnished by the applicant; independently. CSIR-IIP reserves the right to reject any prospective application without assigning any reason.

Brief Details of the GC Analyzer for Refinery Gases (RGA) are as follows.

The Refinery Gas analyser (RGA) unit would include the following:

1. Application: Complete analysis of Pyrolysis gaseous products (H₂, N₂, O₂, CO, CO₂, C₁-C₅+ and oxygenates in single online injection and Co and CO₂ in ppm level detection.
2. Minimum detection of each component: 500 PPM or less.
3. Sample loop should be back flushed to remove the contaminants.
4. Automatic data analysis software and control of entire system.
5. Programmable column oven with rapid cooling from 400 to 50°C.
6. Injection temp: 450°C with split and Split less injection.
7. Automatic data analysis software and control of entire system.
8. CO, CO₂ methanizer bypass valve should be automated.


(S.O. Stores & Purchase)